

ANNEX B

RECOMMENDED TOLERANCES FOR WEIGHING EQUIPMENT

B.1. New Equipment

New weighing machines shall be adjusted and calibrated such that the error of indication, if any, does not exceed 0.25% of the applied load, or 1 scale interval, whichever is the greater.

B.2. Repaired Equipment

Weighing equipment that has been the subject of repair by the replacement, adjustment or modification of any metrologically significant component shall be adjusted and calibrated as if it were new equipment.

B.3. Equipment in Service

Equipment which is in service shall be calibrated to 0.5% of the applied load or 1 scale interval, whichever is the greater.